

AMENDMENT OF SOLICITATION/MODIFICATION OF CONTRACT			1. CONTRACT ID CODE		PAGE OF PAGES 1 2		
2. AMENDMENT/MODIFICATION NO. 0001		3. EFFECTIVE DATE 08/31/2026		4. REQUISITION/PURCHASE REQ. NO.		5. PROJECT NO. (If applicable)	
6. ISSUED BY Aberdeen Area Indian Health Servic Federal Building 115 Fourth Ave, SE Aberdeen SD 57401		CODE 159		7. ADMINISTERED BY (If other than Item 6)		CODE	
8. NAME AND ADDRESS OF CONTRACTOR (No., street, county, State and ZIP Code)				(x) 9A. AMENDMENT OF SOLICITATION NO. 75H70626Q00184			
				x 9B. DATED (SEE ITEM 11) 08/31/2026			
				10A. MODIFICATION OF CONTRACT/ORDER NO.			
				10B. DATED (SEE ITEM 13)			
CODE		FACILITY CODE					

11. THIS ITEM ONLY APPLIES TO AMENDMENTS OF SOLICITATIONS

- ☒ The above numbered solicitation is amended as set forth in Item 14. The hour and date specified for receipt of Offers ☐ is extended. ☒ is not extended.
- Offers must acknowledge receipt of this amendment prior to the hour and date specified in the solicitation or as amended, by one of the following methods: (a) By completing Items 8 and 15, and returning _____ copies of the amendment; (b) By acknowledging receipt of this amendment on each copy of the offer submitted; or (c) By separate letter or electronic communication which includes a reference to the solicitation and amendment numbers. FAILURE OF YOUR ACKNOWLEDGEMENT TO BE RECEIVED AT THE PLACE DESIGNATED FOR THE RECEIPT OF OFFERS PRIOR TO THE HOUR AND DATE SPECIFIED MAY RESULT IN REJECTION OF YOUR OFFER. If by virtue of this amendment you desire to change an offer already submitted, such change may be made by letter or electronic communication, provided each letter or electronic communication makes reference to the solicitation and this amendment, and is received prior to the opening hour and date specified.
12. ACCOUNTING AND APPROPRIATION DATA (If required)

13. THIS ITEM ONLY APPLIES TO MODIFICATION OF CONTRACTS/ORDERS. IT MODIFIES THE CONTRACT/ORDER NO. AS DESCRIBED IN ITEM 14.

CHECK ONE	A. THIS CHANGE ORDER IS ISSUED PURSUANT TO: (Specify authority) THE CHANGES SET FORTH IN ITEM 14 ARE MADE IN THE CONTRACT ORDER NO. IN ITEM 10A.
	B. THE ABOVE NUMBERED CONTRACT/ORDER IS MODIFIED TO REFLECT THE ADMINISTRATIVE CHANGES (such as changes in paying office, appropriation data, etc.) SET FORTH IN ITEM 14, PURSUANT TO THE AUTHORITY OF FAR 43.103(b).
	C. THIS SUPPLEMENTAL AGREEMENT IS ENTERED INTO PURSUANT TO AUTHORITY OF:
	D. OTHER (Specify type of modification and authority)

E. IMPORTANT: Contractor ☐ is not ☐ is required to sign this document and return _____ copies to the issuing office.

14. DESCRIPTION OF AMENDMENT/MODIFICATION (Organized by UCF section headings, including solicitation/contract subject matter where feasible.)

Questions received in reference to 75H70626Q00184 have been answered in the attached document and prompted this amendment to the solicitation.

All other terms, provisions and clauses remain unchanged from the original.
Period of Performance: 09/18/2026 to 08/31/2027

Except as provided herein, all terms and conditions of the document referenced in Item 9 A or 10A, as heretofore changed, remains unchanged and in full force and effect.

15A. NAME AND TITLE OF SIGNER (Type or print)		16A. NAME AND TITLE OF CONTRACTING OFFICER (Type or print) NICOLE R. LEREW	
15B. CONTRACTOR/OFFEROR _____ (Signature of person authorized to sign)	15C. DATE SIGNED	16B. UNITED STATES OF AMERICA _____ (Signature of Contracting Officer)	16C. DATE SIGNED

Previous edition unusable

1. The solicitation calls for Cost-Per-Test (CPT) pricing. Would the Government also consider a fixed monthly fee as an alternative for laboratories where CPT invoicing may be unnecessarily burdensome? **Yes, we will accept a fixed monthly fee as an alternative.**
2. If an analyzer has been in service for less than 12 months when the contract is awarded, must the contractor replace it, or may the Government laboratory and Contracting Officer allow it to remain in place based on its condition and performance? **Yes, we will allow vendor to keep the current analyzers in place if they are of the same model as proposed.**
3. The solicitation indicates that reticulocyte testing and body fluid analysis may be priced separately, but Appendix A does not provide estimated quantities. Will the Government issue a revised pricing schedule if the scope of the requirement occurs? Do proposed solutions simply require the ability to be offer this FDA approved testing capability, should it be required in the future? **The proposed solutions need to have the ability to offer this FDA approved capability.**
4. Section 8 of the Combined Synopsis/Solicitation states that this procurement is a 100% small-business set-aside under NAICS 334516, with a 1,000-employee size standard. Is the set-aside firm, or could a large or other-than-small business be considered for award? **Large other-than-small business CANNOT be considered for this solicitation.**
5. Does your lab have space requirements and limitations that proposed equipment must meet? **Yes. Space limitations vary by testing location. Clinic laboratories have countertops approximately 24 inches deep, while hospital laboratories have countertops approximately 32 inches deep. Proposed equipment must fit within the applicable countertop depth while accommodating all manufacturer-required clearances for ventilation, operation, maintenance, and service access. Vendors should provide the instrument dimensions (W × D × H), required clearances, and total installed footprint, including any space required for external reagents, waste containers, computers, or other accessories. Compact configurations that minimize overall bench-space requirements are preferred.**
6. Does your lab have minimum linearity range requirements for results? **Minimum required low-end performance: WBC: $=0.2 \times 10^3/\mu\text{L}$; RBC: $=0.4 \times 10^6/\mu\text{L}$; PLT: $=8 \times 10^3/\mu\text{L}$**